



TEST REPORT No. 14665.10-19

issued by the Calibration Laboratory of Institut for testing and certification.

Object identification:

Name: One-chamber auto injector
Type: AI1 - T
Manufacturer: OLD AND RICH spol. s r.o.
Serial /Registration No.: 5 pcs of type AI2-OA, the samples have no identification by serial numbers

Customer:

OLD&RICH spol. s r.o.
Drtinova 343/18, Smíchov, 150 00 Praha

Conditions of test:

Ambient temperature: 23 °C ± 5 °C
Relative humidity: 45 % ± 25 %

Workplace of laboratory:

Mesit area, Sokolovská 573, 686 01 Uherské Hradiště

The test was executed on 6th May 2019. It was performed according to the customers' request:

Falling test – For each sample was executed free fall onto the concrete floor. The height of free fall was 1,5 m, number of repeating 5. The origin positions of the sample were different for each of the 5 repeating (horizontally, vertically – needle down, vertically – needle up, askew - needle down, askew - needle up),


Type A II-T passed.

Specifications of Used Standards and Other Measuring Instruments:


Instrument Name	Instrument Type	Serial Number	Calibrated by
Digital standard meter	(0 - 5,5) m	200732	ČMI
Thermometer - hygrometer – data logger		THZ1int Comet	9950034 AKL 2222

All the used standards and measuring instruments are traceable to the national standards of appropriate quantities either directly or through the mediation of verified reference standards.

Uherské Hradiště, 20th May 2019.


Ing. Miroslav Netopil, František Trchalík
Workers Authorized for Execution of measurement




Ing. Miroslav Netopil
Head of Calibration Laboratory